

Unipolar Head (JRI)/Furlong LOL Unipolar Modular Hip Investigation

Note: This analysis compares the Unipolar Head (JRI)/Furlong LOL head/femoral stem combination with all other unipolar modular hip prostheses.

This combination has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, <https://aoanjrr.sahmri.com/annual-reports-2026>.

Note: Procedures using prostheses with no recorded use in 2025 are excluded from the comparator.

TABLE 1

Revision Rate of Primary Unipolar Modular Hip Replacement

The revision rate of the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination is compared to all other unipolar modular hip prostheses.

Table 1: Revision Rates of Primary Unipolar Modular Hip Replacement

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
Unipolar Head (JRI)/Furlong LOL	11	132	565	1.95 (0.97, 3.48)
Other Unipolar Modular Hip	1688	52977	179751	0.94 (0.89, 0.98)
TOTAL	1699	53109	180316	0.94 (0.90, 0.99)

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 2

Yearly Cumulative Percent Revision of Primary Unipolar Modular Hip Replacement

The yearly cumulative percent revision of the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination is compared to all other unipolar modular hip prostheses.

Table 2: Yearly Cumulative Percent Revision (95% CI) of Primary Unipolar Modular Hip Replacement

CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs
Unipolar Head (JRI)/Furlong LOL	6.4 (3.1, 13.0)	9.8 (5.3, 17.5)	9.8 (5.3, 17.5)	11.2 (6.3, 19.6)	11.2 (6.3, 19.6)		
Other Unipolar Modular Hip	2.0 (1.9, 2.2)	2.6 (2.5, 2.8)	3.1 (3.0, 3.3)	3.7 (3.5, 3.9)	4.2 (4.0, 4.4)	4.7 (4.4, 4.9)	5.2 (4.9, 5.5)

CPR	8 Yrs	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs
Unipolar Head (JRI)/Furlong LOL							
Other Unipolar Modular Hip	5.7 (5.4, 6.0)	6.2 (5.8, 6.6)	6.7 (6.3, 7.1)	7.1 (6.6, 7.6)	7.5 (6.9, 8.0)	7.9 (7.3, 8.6)	8.5 (7.8, 9.3)

CPR	15 Yrs	16 Yrs	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs
Unipolar Head (JRI)/Furlong LOL							
Other Unipolar Modular Hip	8.7 (8.0, 9.5)	9.5 (8.5, 10.6)	9.5 (8.5, 10.6)	10.1 (8.9, 11.6)	10.6 (9.1, 12.3)	11.2 (9.3, 13.5)	11.2 (9.3, 13.5)

Note: Prostheses no longer used in 2025 are excluded from the comparator.


FIGURE 1

Yearly Cumulative Percent Revision of Primary Unipolar Modular Hip Replacement

The yearly cumulative percent revision of the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination is compared to all other unipolar modular hip prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

Figure 1: Cumulative Percent Revision of Primary Unipolar Modular Hip Replacement



Number at Risk	0 Yr	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs
Unipolar Head (JRI)/Furlong LOL	132	89	74	63	56	42	36	31
Other Unipolar Modular Hip	52977	35569	28270	22155	17016	13156	10100	7733

Number at Risk	8 Yrs	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
Unipolar Head (JRI)/Furlong LOL	29	24	19	14	11	7	6	5	4
Other Unipolar Modular Hip	5914	4507	3394	2585	1981	1484	1109	783	537

Number at Risk	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs	24 Yrs	25 Yrs
Unipolar Head (JRI)/Furlong LOL	4	1	0	0	0	0	0	0	0
Other Unipolar Modular Hip	366	245	145	90	54	29	11	4	1

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 3**Primary Diagnosis for Revised Primary Unipolar Modular Hip Replacement**

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other unipolar modular hip prostheses.

Table 3: Primary Diagnosis for Revised Primary Unipolar Modular Hip Replacement

Primary Diagnosis	Unipolar Head (JRI)/Furlong LOL		Other Unipolar Modular Hip	
	Number	Percent	Number	Percent
Fractured Neck Of Femur	11	100.0	1601	94.8
Osteoarthritis			46	2.7
Failed Internal Fixation			25	1.5
Osteonecrosis			8	0.5
Tumour			6	0.4
Other			1	0.1
Rheumatoid Arthritis			1	0.1
TOTAL	11	100.0	1688	100.0

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 4

Reasons for Revision

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

% Primaries Revised: This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

% Revisions: The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

Table 4: Primary Unipolar Modular Hip Replacement - Reason for Revision (Follow-up Limited to 18.4 Years)

Revision Diagnosis	Unipolar Head (JRI)/Furlong LOL			Other Unipolar Modular Hip		
	Number	% Primaries Revised	% Revisions	Number	% Primaries Revised	% Revisions
Infection	3	2.3	27.3	439	0.8	26.0
Prosthesis Dislocation/Instability	2	1.5	18.2	361	0.7	21.4
Chondrolysis/Acetab. Erosion				267	0.5	15.8
Fracture	2	1.5	18.2	264	0.5	15.6
Pain	1	0.8	9.1	163	0.3	9.7
Loosening	3	2.3	27.3	132	0.2	7.8
Lysis				21	0.0	1.2
Implant Breakage Stem				7	0.0	0.4
Incorrect Sizing				6	0.0	0.4
Leg Length Discrepancy				6	0.0	0.4
Metal Related Pathology				5	0.0	0.3
Malposition				3	0.0	0.2
Tumour				3	0.0	0.2
Osteonecrosis				1	0.0	0.1
Other				9	0.0	0.5
N Revision	11	8.3	100.0	1687	3.2	100.0
N Primary	132			52977		

Note: This table is restricted to revisions within 18.4 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2025 are excluded from the comparator.

FIGURE 2**Cumulative Incidence Revision Diagnosis of Primary Unipolar Modular Hip Replacement**

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other unipolar modular hip prostheses.


Figure 2: Cumulative Incidence Revision Diagnosis for Primary Unipolar Modular Hip Replacement

TABLE 5

Type of Revision Performed for Primary Unipolar Modular Hip Replacement

This analysis identifies the components used in the revision of the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination and compares it to the components used in the revision of all other unipolar modular hip prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other unipolar modular hip prostheses i.e. is there a difference in the type of revision undertaken for the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination compared to all other unipolar modular hip prostheses.

Table 5: Primary Unipolar Modular Hip Replacement - Type of Revision (Follow-up Limited to 18.4 Years)

Type of Revision	Unipolar Head (JRI)/Furlong LOL		Other Unipolar Modular Hip	
	Number	Percent	Number	Percent
Acetabular Component	2	18.2	654	38.8
THR (Femoral/Acetabular)	3	27.3	283	16.8
Femoral Component	1	9.1	173	10.3
Bipolar Head and Femoral	2	18.2	81	4.8
Removal of Prostheses			59	3.5
Cement Spacer			58	3.4
Reinsertion of Components			7	0.4
N Major	8	72.7	1315	77.9
Head Only	2	18.2	297	17.6
Minor Components	1	9.1	46	2.7
Bipolar Only			27	1.6
Head/Insert			2	0.1
N Minor	3	27.3	372	22.1
TOTAL	11	100.0	1687	100.0

Note: This table is restricted to revisions within 18.4 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 6**Revision Rates of Unipolar Head (JRI)/Furlong LOL Primary Unipolar Modular Hip Replacement by Fixation**

This analysis is provided as some prostheses have more than one fixation option. Additionally there are prostheses where an alternative to the recommended approach to fixation was used e.g. a cementless prosthesis that has been cemented or vice-versa.

Table 6: Revised Number of Unipolar Head (JRI)/Furlong LOL Primary Unipolar Modular Hip Replacement by Fixation

Fixation	N Revised	N Total
Cementless	11	132
TOTAL	11	132

TABLE 7

Revision Rates of Primary Unipolar Modular Hip Replacement by State

This enables a state by state variation to be identified for the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination and provides the comparative data for each of the states for all other unipolar modular hip prostheses.

The purpose of this analysis is to determine if the higher than anticipated rate of revision has widespread distribution between states. If there is widespread distribution then the reason for the higher than anticipated rate of revision is unlikely to be surgeon specific. If the prosthesis has been used in only a small number of states it is not possible to distinguish if the higher than anticipated rate of revision is related to the prosthesis, surgeon, technique or patient.

Table 7: Revised Number of Primary Unipolar Modular Hip Replacement by State

Component	State	N Revised	N Total
Unipolar Head (JRI)/Furlong LOL	NSW	0	2
	VIC	5	86
	QLD	5	42
	SA	1	2
Other Unipolar Modular Hip	NSW	341	12014
	VIC	523	16860
	QLD	282	8821
	WA	257	7496
	SA	244	6407
	TAS	24	1038
	ACT/NT	17	341
TOTAL		1699	53109

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 8**Number of Revisions of Unipolar Head (JRI)/Furlong LOL Primary Unipolar Modular Hip Replacement by Year of Implant**

This analysis details the number of prostheses reported each year to the Registry for the Unipolar Head (JRI)/Furlong LOL unipolar modular hip combination. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2025 has a maximum of one year to be revised, whereas a primary procedure performed in 2023 has a maximum of three years to be revised.

Table 8: Number of Revisions of Unipolar Head (JRI)/Furlong LOL Primary Unipolar Modular Hip Replacement by Year of Implant

Year of Implant	Number Revised	Total Number
2007	1	12
2008	1	18
2009	3	10
2010	1	13
2011	1	10
2012	1	8
2013	1	7
2014	1	34
2015	0	16
2016	1	4
TOTAL	11	132

TABLE 9

Revision Rates of Unipolar Head (JRI)/Furlong LOL Primary Unipolar Modular Hip Replacement by Catalogue Number Range

Many prostheses have a number of catalogue ranges. The catalogue range is specific to particular design features; more than one catalogue range usually indicates a minor difference in design in a particular Unipolar Head (JRI)/Furlong LOL prosthesis.

This analysis has been undertaken to determine if the revision rate varies according to the catalogue number range.

Model	Catalogue Range	Catalogue Description	Cement	Material	Coating
Head					
Unipolar Head (JRI)	934001-935801	STAINLESS STEEL PHYSIOLOGICAL HEAD	NO	METAL	
Femoral Stem					
Furlong LOL	970900-971400	LOL H-AC STEM	NO	METAL	HA COATED

Table 9: Revised Number of Unipolar Head (JRI)/Furlong LOL Primary Unipolar Modular Hip Replacement by Catalogue Number Range

Head Range	Femoral Stem Range	N Revised	N Total
934001-935801	970900-971400	11	132
TOTAL		11	132