

Optetrak Logic PS Total Knee Investigation

Note: This analysis compares the Optetrak Logic PS femoral prosthesis with all other total knee prostheses.

This prosthesis has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, <https://aoanjrr.sahmri.com/annual-reports-2026>.

Note: Procedures using prostheses with no recorded use in 2025 are excluded from the comparator.

TABLE 1

Revision Rate of Primary Total Knee Replacement

The revision rate of the Optetrak Logic PS total knee prosthesis is compared to all other total knee prostheses.

Table 1: Revision Rates of Primary Total Knee Replacement

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
Optetrak Logic PS	159	1751	10930	1.45 (1.24, 1.70)
Other Total Knee	27284	791430	5145922	0.53 (0.52, 0.54)
TOTAL	27443	793181	5156852	0.53 (0.53, 0.54)

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 2

Yearly Cumulative Percent Revision of Primary Total Knee Replacement

The yearly cumulative percent revision of the Optetrak Logic PS total knee prosthesis is compared to all other total knee prostheses.

Table 2: Yearly Cumulative Percent Revision (95% CI) of Primary Total Knee Replacement

CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs
Optetrak Logic PS	1.7 (1.2, 2.4)	2.6 (2.0, 3.5)	3.5 (2.7, 4.5)	4.9 (4.0, 6.0)	6.3 (5.2, 7.6)	7.5 (6.3, 9.0)
Other Total Knee	1.0 (1.0, 1.0)	1.8 (1.8, 1.8)	2.4 (2.3, 2.4)	2.8 (2.7, 2.8)	3.1 (3.1, 3.1)	3.4 (3.4, 3.5)

CPR	7 Yrs	8 Yrs	9 Yrs	10 Yrs	11 Yrs	12 Yrs
Optetrak Logic PS	9.3 (7.8, 11.0)	10.5 (8.9, 12.5)	12.3 (10.3, 14.7)	16.7 (13.6, 20.3)		
Other Total Knee	3.7 (3.7, 3.8)	4.0 (4.0, 4.1)	4.3 (4.3, 4.4)	4.6 (4.6, 4.7)	4.9 (4.9, 5.0)	5.3 (5.2, 5.3)

CPR	13 Yrs	14 Yrs	15 Yrs	16 Yrs	17 Yrs	18 Yrs
Optetrak Logic PS						
Other Total Knee	5.6 (5.5, 5.7)	5.9 (5.8, 6.0)	6.3 (6.2, 6.4)	6.7 (6.6, 6.9)	7.2 (7.0, 7.3)	7.5 (7.4, 7.7)

CPR	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs	24 Yrs	25 Yrs
Optetrak Logic PS							
Other Total Knee	7.9 (7.7, 8.0)	8.1 (7.9, 8.3)	8.5 (8.2, 8.7)	8.7 (8.4, 9.0)	8.8 (8.6, 9.2)	9.1 (8.7, 9.5)	9.5 (8.8, 10.2)

Note: Prostheses no longer used in 2025 are excluded from the comparator.


FIGURE 1

Yearly Cumulative Percent Revision of Primary Total Knee Replacement

The yearly cumulative percent revision of the Optetrak Logic PS total knee prosthesis is compared to all other total knee prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

Figure 1: Cumulative Percent Revision of Primary Total Knee Replacement



Number at Risk	0 Yr	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs
Optetrak Logic PS	1751	1716	1685	1605	1445	1226	919	615
Other Total Knee	791430	708699	628274	550835	487544	424868	369845	315999

Number at Risk	8 Yrs	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
Optetrak Logic PS	424	289	119	1	0	0	0	0	0
Other Total Knee	266816	222320	182654	146980	117201	92187	70380	52526	38351

Number at Risk	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs	24 Yrs	25 Yrs
Optetrak Logic PS	0	0	0	0	0	0	0	0	0
Other Total Knee	27628	19434	13616	9427	6164	3657	2096	972	244

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 3

Primary Diagnosis for Revised Primary Total Knee Replacement

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other total knee prostheses.

Table 3: Primary Diagnosis for Revised Primary Total Knee Replacement

Primary Diagnosis	Optetrak Logic PS		Other Total Knee	
	Number	Percent	Number	Percent
Osteoarthritis	158	99.4	26392	96.7
Rheumatoid Arthritis			330	1.2
Tumour			199	0.7
Other Inflammatory Arthritis	1	0.6	177	0.6
Osteonecrosis			104	0.4
Fracture			55	0.2
Other			26	0.1
Chondrocalcinosis			1	0.0
TOTAL	159	100.0	27284	100.0

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 4

Reasons for Revision

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

% Primaries Revised: This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

% Revisions: The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

Table 4: Primary Total Knee Replacement - Reason for Revision (Follow-up Limited to 11.1 Years)

Revision Diagnosis	Optetrak Logic PS			Other Total Knee		
	Number	% Primaries Revised	% Revisions	Number	% Primaries Revised	% Revisions
Infection	39	2.2	24.5	7609	1.0	29.9
Loosening	56	3.2	35.2	5308	0.7	20.9
Instability	8	0.5	5.0	2734	0.3	10.8
Patella Erosion	1	0.1	0.6	1778	0.2	7.0
Pain	6	0.3	3.8	1766	0.2	6.9
Patellofemoral Pain				1471	0.2	5.8
Arthrofibrosis	4	0.2	2.5	1168	0.1	4.6
Fracture	4	0.2	2.5	985	0.1	3.9
Malalignment	2	0.1	1.3	503	0.1	2.0
Lysis	9	0.5	5.7	232	0.0	0.9
Wear Tibial Insert	7	0.4	4.4	231	0.0	0.9
Incorrect Sizing	1	0.1	0.6	217	0.0	0.9
Patella Maltracking	2	0.1	1.3	191	0.0	0.8
Implant Breakage Tibial Insert	4	0.2	2.5	165	0.0	0.6
Bearing Dislocation	1	0.1	0.6	142	0.0	0.6
Implant Breakage Patella	2	0.1	1.3	138	0.0	0.5
Metal Related Pathology	2	0.1	1.3	106	0.0	0.4
Prosthesis Dislocation				75	0.0	0.3
Synovitis				62	0.0	0.2
Osteonecrosis				50	0.0	0.2
Wear Patella	2	0.1	1.3	41	0.0	0.2
Implant Breakage Femoral	1	0.1	0.6	36	0.0	0.1
Implant Breakage Tibial				36	0.0	0.1
Tumour				33	0.0	0.1
Heterotopic Bone				14	0.0	0.1
Progression Of Disease				7	0.0	0.0
Wear Tibial				5	0.0	0.0
Patella Dislocation				3	0.0	0.0
Incorrect Side				1	0.0	0.0
Wear Femoral				1	0.0	0.0
Other	8	0.5	5.0	308	0.0	1.2
N Revision	159	9.1	100.0	25416	3.2	100.0
N Primary	1751			791430		

Note: This table is restricted to revisions within 11.1 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2025 are excluded from the comparator.

FIGURE 2

Cumulative Incidence Revision Diagnosis of Primary Total Knee Replacement

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the Optetrak Logic PS total knee prosthesis. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other total knee prostheses.

Figure 2: Cumulative Incidence Revision Diagnosis for Primary Total Knee Replacement




TABLE 5

Type of Revision Performed for Primary Total Knee Replacement

This analysis identifies the components used in the revision of the Optetrak Logic PS total knee prosthesis and compares it to the components used in the revision of all other total knee prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other total knee prostheses i.e. is there a difference in the type of revision undertaken for the Optetrak Logic PS total knee prosthesis compared to all other total knee prostheses.

Table 5: Primary Total Knee Replacement - Type of Revision (Follow-up Limited to 11.1 Years)

Type of Revision	Optetrak Logic PS		Other Total Knee	
	Number	Percent	Number	Percent
TKR (Tibial/Femoral)	66	41.5	6008	23.6
Tibial Component	2	1.3	1859	7.3
Femoral Component	8	5.0	1275	5.0
Cement Spacer	4	2.5	1103	4.3
Removal of Prostheses	3	1.9	130	0.5
Total Femoral			20	0.1
Reinsertion of Components			8	0.0
N Major	83	52.2	10403	40.9
Insert Only	33	20.8	8210	32.3
Patella Only	8	5.0	4285	16.9
Insert/Patella	35	22.0	2445	9.6
Minor Components			64	0.3
Cement Only			9	0.0
N Minor	76	47.8	15013	59.1
TOTAL	159	100.0	25416	100.0

Note: This table is restricted to revisions within 11.1 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 6**Revision Rates of Optetrak Logic PS Primary Total Knee Replacement by Fixation**

This analysis is provided as some prostheses have more than one fixation option. Additionally there are prostheses where an alternative to the recommended approach to fixation was used e.g. a cementless prosthesis that has been cemented or vice-versa.

Table 6: Revised Number of Optetrak Logic PS Primary Total Knee Replacement by Fixation

Fixation	N Revised	N Total
Cemented	159	1749
Hybrid (Tibial Cemented)	0	2
TOTAL	159	1751

TABLE 7**Revision Rates of Optetrak Logic PS Primary Total Knee Replacement by Bearing Surface**

This analysis is provided as some prostheses are combined with a variety of bearing surfaces. All bearing surfaces used with this prosthesis are listed.

Table 7: Revised Number of Optetrak Logic PS Primary Total Knee Replacement by Bearing Surface

Bearing Surface	N Revised	N Total
Non XLPE	159	1751
TOTAL	159	1751

TABLE 8**Revision Rates of Optetrak Logic PS Primary Total Knee Replacement by Bearing Mobility**

This analysis is provided as some prostheses are combined with a variety of bearing mobilities. All bearing mobilities used with this prosthesis are listed.

Table 8: Revised Number of Optetrak Logic PS Primary Total Knee Replacement by Bearing Mobility

Bearing Mobility	N Revised	N Total
Fixed	76	656
Rotating	83	1095
TOTAL	159	1751

TABLE 9**Revision Rates of Optetrak Logic PS Primary Total Knee Replacement by Stability**

This analysis is provided as some prostheses are combined with a variety of stabilities. All stabilities used with this prosthesis are listed.

Table 9: Revised Number of Optetrak Logic PS Primary Total Knee Replacement by Stability

Stability	N Revised	N Total
Posterior Stabilised	159	1751
TOTAL	159	1751

TABLE 10

Revision Rates of Primary Total Knee Replacement by State

This enables a state by state variation to be identified for the Optetrak Logic PS total knee prosthesis and provides the comparative data for each of the states for all other total knee prostheses.

The purpose of this analysis is to determine if the higher than anticipated rate of revision has widespread distribution between states. If there is widespread distribution then the reason for the higher than anticipated rate of revision is unlikely to be surgeon specific. If the prosthesis has been used in only a small number of states it is not possible to distinguish if the higher than anticipated rate of revision is related to the prosthesis, surgeon, technique or patient.

Table 10: Revised Number of Primary Total Knee Replacement by State

Component	State	N Revised	N Total
Optetrak Logic PS	NSW	64	803
	VIC	1	21
	QLD	77	741
	WA	0	2
	TAS	17	184
Other Total Knee	NSW	7550	259497
	VIC	6020	163565
	QLD	5789	167244
	WA	3600	92841
	SA	3220	73114
	TAS	475	15075
	ACT/NT	630	20094
TOTAL		27443	793181

Note: Prostheses no longer used in 2025 are excluded from the comparator.

TABLE 11**Number of Revisions of Optetrak Logic PS Primary Total Knee Replacement by Year of Implant**

This analysis details the number of prostheses reported each year to the Registry for the Optetrak Logic PS total knee prosthesis. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2025 has a maximum of one year to be revised, whereas a primary procedure performed in 2023 has a maximum of three years to be revised.

Table 11: Number of Revisions of Optetrak Logic PS Primary Total Knee Replacement by Year of Implant

Year of Implant	Number Revised	Total Number
2014	1	2
2015	19	172
2016	23	203
2017	11	163
2018	17	211
2019	30	313
2020	35	324
2021	14	188
2022	8	130
2023	1	45
TOTAL	159	1751

TABLE 12

Revision Rates of Optetrak Logic PS Primary Total Knee Replacement by Catalogue Number Range

Many prostheses have a number of catalogue ranges. The catalogue range is specific to particular design features; more than one catalogue range usually indicates a minor difference in design in a particular Optetrak Logic PS prosthesis.

This analysis has been undertaken to determine if the revision rate varies according to the catalogue number range.

Model	Catalogue Range	Catalogue Description	Cement
Femoral			
Optetrak Logic PS	02010010200-02010010360	OPTETRAK LOGIC FEMORAL COMPONENT PS CEMENTED	YES
Optetrak Logic PS	02010020200-02010020360	OPTETRAK LOGIC FEMORAL COMPONENT PS POROUS	NO

Table 12: Revised Number of Optetrak Logic PS Primary Total Knee Replacement by Catalogue Number Range

Femoral Range	N Revised	N Total
02010010200-02010010360	159	1749
02010020200-02010020360	0	2
TOTAL	159	1751

TABLE 13**Revision Rates of Optetrak Logic PS Primary Total Knee Replacement by Component**

A prosthesis may be combined with multiple components. This analysis has been undertaken to determine if the revision rate varies according to the component with which it is combined.

Table 13: Revised Number of Optetrak Logic PS Primary Total Knee Replacement by Tibial Component

Tibial Component	N Revised	N Total
Optetrak Logic	75	652
Optetrak Logic RBK	84	1099
TOTAL	159	1751