Cormet Total Conventional Hip Investigation

Note: This analysis compares the Cormet acetabular prosthesis with all other total conventional hip prostheses.

This prosthesis has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, https://aoanjrr.sahmri.com/annual-reports-2026.

Note: Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator. Procedures using prostheses with no recorded use in 2025 are excluded from the comparator.

TABLE 1

Revision Rate of Primary Total Conventional Hip Replacement

The revision rate of the Cormet total conventional hip prosthesis is compared to all other total conventional hip prostheses.

Table 1: Revision Rates of Primary Total Conventional Hip Replacement

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
Cormet	167	803	10603	1.58 (1.35, 1.83)
Other Total Conventional Hip	19701	568612	3623071	0.54 (0.54, 0.55)
TOTAL	**19868**	**569415**	**3633674**	**0.55 (0.54, 0.55)**

Note: Prostheses no longer used in 2025 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 2

Yearly Cumulative Percent Revision of Primary Total Conventional Hip Replacement

The yearly cumulative percent revision of the Cormet total conventional hip prosthesis is compared to all other total conventional hip prostheses.

Table 2: Yearly Cumulative Percent Revision (95% CI) of Primary Total Conventional Hip Replacement

CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs
Cormet	1.5 (0.9, 2.6)	2.4 (1.5, 3.7)	3.5 (2.4, 5.1)	4.2 (3.0, 5.8)	5.2 (3.9, 7.0)	6.9 (5.4, 9.0)
Other Total Conventional Hip	1.7 (1.7, 1.8)	2.1 (2.1, 2.2)	2.5 (2.4, 2.5)	2.8 (2.7, 2.8)	3.0 (3.0, 3.1)	3.3 (3.2, 3.3)

CPR	7 Yrs	8 Yrs	9 Yrs	10 Yrs	11 Yrs	12 Yrs
Cormet	9.0 (7.2, 11.3)	10.3 (8.3, 12.7)	12.2 (10.1, 14.8)	13.7 (11.4, 16.5)	15.2 (12.7, 18.0)	16.1 (13.6, 19.0)
Other Total Conventional Hip	3.5 (3.5, 3.6)	3.8 (3.7, 3.9)	4.1 (4.0, 4.2)	4.4 (4.3, 4.5)	4.7 (4.6, 4.8)	5.1 (5.0, 5.2)

CPR	13 Yrs	14 Yrs	15 Yrs	16 Yrs	17 Yrs	18 Yrs
Cormet	18.0 (15.3, 21.1)	19.6 (16.8, 22.8)	20.8 (17.9, 24.1)	22.3 (19.3, 25.7)	24.4 (21.2, 28.1)	28.1 (24.2, 32.3)
Other Total Conventional Hip	5.4 (5.3, 5.5)	5.8 (5.7, 5.9)	6.2 (6.1, 6.3)	6.6 (6.5, 6.7)	7.0 (6.8, 7.1)	7.3 (7.1, 7.5)

CPR	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs	24 Yrs	25 Yrs
Cormet	28.1 (24.2, 32.3)	28.1 (24.2, 32.3)	28.1 (24.2, 32.3)				
Other Total Conventional Hip	7.7 (7.5, 7.9)	8.0 (7.8, 8.2)	8.6 (8.3, 8.9)	9.0 (8.7, 9.4)	9.6 (9.1, 10.0)	9.6 (9.2, 10.1)	10.3 (9.4, 11.2)

Note: Prostheses no longer used in 2025 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

FIGURE 1

Yearly Cumulative Percent Revision of Primary Total Conventional Hip Replacement

The yearly cumulative percent revision of the Cormet total conventional hip prosthesis is compared to all other total conventional hip prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

Figure 1: Cumulative Percent Revision of Primary Total Conventional Hip Replacement



Number at Risk	0 Yr	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs
Cormet	803	787	775	754	739	718	685	644
Other Total Conventional Hip	568612	501603	445433	391564	344389	299354	259516	220948

Number at Risk	8 Yrs	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
Cormet	617	587	560	531	506	477	452	409	332
Other Total Conventional Hip	185798	153739	124848	99575	79802	63343	49730	38033	28355

Number at Risk	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs	24 Yrs	25 Yrs
Cormet	230	147	109	70	44	23	4	1	0
Other Total Conventional Hip	20437	14638	10400	7396	4859	2863	1455	560	95

Note: Prostheses no longer used in 2025 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 3

Primary Diagnosis for Revised Primary Total Conventional Hip Replacement

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other total conventional hip prostheses.

Table 3: Primary Diagnosis for Revised Primary Total Conventional Hip Replacement

	Cormet		Other Total Conventional Hip	
Primary Diagnosis	Number	Percent	Number	Percent
Osteoarthritis	145	86.8	16296	82.7
Fractured Neck Of Femur	7	4.2	1521	7.7
Osteonecrosis	6	3.6	844	4.3
Developmental Dysplasia	2	1.2	321	1.6
Rheumatoid Arthritis	1	0.6	201	1.0
Failed Internal Fixation			165	0.8
Tumour	1	0.6	157	0.8
Other Inflammatory Arthritis	4	2.4	102	0.5
Fracture/Dislocation			60	0.3
Other	1	0.6	21	0.1
Arthrodesis Takedown			13	0.1
TOTAL	**167**	**100.0**	**19701**	**100.0**

Note: Prostheses no longer used in 2025 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 4

Reasons for Revision

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

% Primaries Revised: This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

% Revisions: The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

Table 4: Primary Total Conventional Hip Replacement - Reason for Revision

	Cormet			Other Total Conventional Hip		
Revision Diagnosis	Number	% Primaries Revised	% Revisions	Number	% Primaries Revised	% Revisions
Infection	27	3.4	16.2	5017	0.9	25.5
Prosthesis Dislocation/Instability	6	0.7	3.6	4382	0.8	22.2
Fracture	25	3.1	15.0	4349	0.8	22.1
Loosening	41	5.1	24.6	3629	0.6	18.4
Pain	4	0.5	2.4	340	0.1	1.7
Leg Length Discrepancy	3	0.4	1.8	307	0.1	1.6
Malposition				282	0.0	1.4
Implant Breakage Stem	11	1.4	6.6	211	0.0	1.1
Lysis	9	1.1	5.4	200	0.0	1.0
Implant Breakage Acetabular Insert				128	0.0	0.6
Wear Acetabular Insert				112	0.0	0.6
Incorrect Sizing	1	0.1	0.6	96	0.0	0.5
Metal Related Pathology	36	4.5	21.6	93	0.0	0.5
Implant Breakage Acetabular	2	0.2	1.2	73	0.0	0.4
Wear Head				49	0.0	0.2
Tumour				46	0.0	0.2
Heterotopic Bone				26	0.0	0.1
Implant Breakage Head				26	0.0	0.1
Wear Acetabulum	1	0.1	0.6	6	0.0	0.0
Osteonecrosis	1	0.1	0.6	4	0.0	0.0
Synovitis				2	0.0	0.0
Progression Of Disease				1	0.0	0.0
Other				322	0.1	1.6
N Revision	**167**	**20.8**	**100.0**	**19701**	**3.5**	**100.0**
N Primary	**803**			**568612**		

Note: Prostheses no longer used in 2025 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

FIGURE 2

Cumulative Incidence Revision Diagnosis of Primary Total Conventional Hip Replacement

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the Cormet total conventional hip prosthesis. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other total conventional hip prostheses.

Figure 2: Cumulative Incidence Revision Diagnosis for Primary Total Conventional Hip Replacement



TABLE 5

Type of Revision Performed for Primary Total Conventional Hip Replacement

This analysis identifies the components used in the revision of the Cormet total conventional hip prosthesis and compares it to the components used in the revision of all other total conventional hip prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other total conventional hip prostheses i.e. is there a difference in the type of revision undertaken for the Cormet total conventional hip prosthesis compared to all other total conventional hip prostheses.

Table 5: Primary Total Conventional Hip Replacement - Type of Revision

	Cormet		Other Total Conventional Hip	
Type of Revision	Number	Percent	Number	Percent
Femoral Component	26	15.6	6645	33.7
Acetabular Component	66	39.5	3311	16.8
THR (Femoral/Acetabular)	51	30.5	2305	11.7
Cement Spacer	15	9.0	556	2.8
Removal of Prostheses			93	0.5
Reinsertion of Components			27	0.1
Total Femoral			12	0.1
Bipolar Head and Femoral	1	0.6	10	0.1
N Major	**159**	**95.2**	**12959**	**65.8**
Head/Insert	5	3.0	5372	27.3
Head Only	2	1.2	914	4.6
Minor Components	1	0.6	285	1.4
Insert Only			168	0.9
Bipolar Only			1	0.0
Cement Only			1	0.0
Head/Neck/Insert			1	0.0
N Minor	**8**	**4.8**	**6742**	**34.2**
TOTAL	**167**	**100.0**	**19701**	**100.0**

Note: Prostheses no longer used in 2025 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 6

Revision Rates of Cormet Primary Total Conventional Hip Replacement by Fixation

This analysis is provided as some prostheses have more than one fixation option. Additionally there are prostheses where an alternative to the recommended approach to fixation was used e.g. a cementless prosthesis that has been cemented or vice-versa.

Table 6: Revised Number of Cormet Primary Total Conventional Hip Replacement by Fixation

Fixation	N Revised	N Total
Cemented	0	1
Cementless	128	629
Hybrid (Femur Cemented)	39	173
TOTAL	**167**	**803**

TABLE 7

Revision Rates of Cormet Primary Total Conventional Hip Replacement by Bearing Surface

This analysis is provided as some prostheses are combined with a variety of bearing surfaces. All bearing surfaces used with this prosthesis are listed.

Table 7: Revised Number of Cormet Primary Total Conventional Hip Replacement by Bearing Surface

Bearing Surface	N Revised	N Total
Metal/Metal	167	803
TOTAL	**167**	**803**

TABLE 8

Revision Rates of Primary Total Conventional Hip Replacement by State

This enables a state by state variation to be identified for the Cormet total conventional hip prosthesis and provides the comparative data for each of the states for all other total conventional hip prostheses.

The purpose of this analysis is to determine if the higher than anticipated rate of revision has widespread distribution between states. If there is widespread distribution then the reason for the higher than anticipated rate of revision is unlikely to be surgeon specific. If the prosthesis has been used in only a small number of states it is not possible to distinguish if the higher than anticipated rate of revision is related to the prosthesis, surgeon, technique or patient.

Table 8: Revised Number of Primary Total Conventional Hip Replacement by State

Component	State	N Revised	N Total
Cormet	NSW	97	508
	VIC	16	49
	QLD	29	106
	SA	4	16
	TAS	21	124
Other Total Conventional Hip	NSW	5507	167002
	VIC	4853	146558
	QLD	3968	104359
	WA	2510	64946
	SA	1887	52084
	TAS	439	18269
	ACT/NT	537	15394
TOTAL		**19868**	**569415**

Note: Prostheses no longer used in 2025 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 9

Number of Revisions of Cormet Primary Total Conventional Hip Replacement by Year of Implant

This analysis details the number of prostheses reported each year to the Registry for the Cormet total conventional hip prosthesis. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2025 has a maximum of one year to be revised, whereas a primary procedure performed in 2023 has a maximum of three years to be revised.

Table 9: Number of Revisions of Cormet Primary Total Conventional Hip Replacement by Year of Implant

Year of Implant	Number Revised	Total Number
2001	0	1
2002	1	8
2003	12	53
2004	21	74
2005	37	103
2006	24	114
2007	15	73
2008	32	129
2009	12	124
2010	10	93
2011	3	26
2012	0	4
2013	0	1
TOTAL	**167**	**803**

TABLE 10

Revision Rates of Cormet Primary Total Conventional Hip Replacement by Catalogue Number Range

Many prostheses have a number of catalogue ranges. The catalogue range is specific to particular design features; more than one catalogue range usually indicates a minor difference in design in a particular Cormet prosthesis.

This analysis has been undertaken to determine if the revision rate varies according to the catalogue number range.

Model	Catalogue Range	Catalogue Description	Cement	Material
Acetabular				
Cormet	179242B-179264B	MOM HAP BI COATED PEGLESS RESURFACING CUP	NO	METAL
Cormet	179544B-179564B	MOM HAP BI COATED DYSPLASIA RESURFACING CUP	NO	METAL
Cormet	479248B-479262B	COCR HAP BI COATED PEGLESS RESURFACING CUP	NO	METAL

Table 10: Revised Number of Cormet Primary Total Conventional Hip Replacement by Catalogue Number Range

Acetabular Range	N Revised	N Total
179242B-179264B	165	784
179544B-179564B	1	18
479248B-479262B	1	1
TOTAL	**167**	**803**

TABLE 11

Revision Rates of Cormet Primary Total Conventional Hip Replacement by Component

A prosthesis may be combined with multiple components. This analysis has been undertaken to determine if the revision rate varies according to the component with which it is combined.

Table 11: Revised Number of Cormet Primary Total Conventional Hip Replacement by Femoral Stem Component

Femoral Stem Component	N Revised	N Total
Accolade I	14	25
Acor	29	225
Alpha Fit	0	4
Apex	0	4
CORAIL	0	1
Custom Made (MSA)	0	1
Emperion	0	2
Exeter V40	0	1
Freeman	4	12
Integrale (exch neck)	6	10
Lubinus SP II	1	1
Metafix	11	47
Mets	0	1
Proxima	3	14
S-Rom	5	41
SL-Plus	45	202
SLR	0	1
Stability	2	5
Summit	0	1
Synergy	0	1
Taper Fit	32	157
Tri-Fit	15	47
TOTAL	**167**	**803**